



GOVERNMENT OF KERALA
GOVERNMENT MEDICAL COLLEGE KOTTAYAM
GANDHINAGAR P.O KOTTAYAM-08

Email ID –
medicalcollegekottayam@gmail.com
Fax- 0481-2597284

Phone Office 0481-2597279
0481-2597284

No.GMCKTM/3195/2025-C5

Date:06-12-2025

SHORT QUOTATION NOTICE

Sealed quotations are invited for the supply of following stores:
"OTOSCOPE -1 NO for the Department of Pediatrics"

The envelopes containing the quotations should bear the superscription "QUOTATION FOR THE PURCHASE OF OTOSCOPE -1 NO for the department of Pediatrics addressed to "THE PRINCIPAL, GOVT. MEDICAL COLLEGE KOTTAYAM-8". Intending tenderers may submit the quotations in their own papers.

Last date and time for the receipt of quotations is **12.12.2025 12 pm**. Late quotations will not be accepted. The quotations will be opened at **12.12.2025, 2 pm**, the presence of such of the tenderers or their authorized representatives who may be present at that time. The maximum period required for delivery of the articles should be mentioned.

Details of the requirements and the conditions governing their supply can be obtained free on request from the Principal, Govt Medical College Kottayam till **12.12.2025, 12 noon**.

Specification of the item is enclosed herewith, for further details , contact with the HOD Paediatrics.

Signed by

Dr Varghese P Punnoose

PRINCIPAL

Date: 06-12-2025 05:42:30

Copy To :Librarian Central Library for publishing in the website, Notice Boards MCK,
MCH, ICH, DCK, CNK, OC/SF
HOD Paediatrics

Dept. of Pediatrics
ICH, GOVT. MEDICAL COLLEGE
KOTTAYAM

CS

7/11/25

Technical Specification

Equipment :Otoscope

1. Should be a C-Cell type or D-cell type Otoscope suitable for usage in outpatient departments of Govt. Hospitals.
2. Should be provided with a halogen/LED light source. In case of halogen light source, 2 Nos of spare bulbs are to be provided.
3. Should be able to detach the otoscope head.
4. Should provide no reflections and obstructions.
5. Should provide the following detachable accessories
 1. Specula (2.5mm to 5mm, 4 sizes)
6. Should have rechargeable battery. Charger for the Rechargeable battery shall be provided.
7. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid

[Signature]
Dr. Jayaprakash
K.P.

