GOVERNMENT OF KERALAM
GOVERNMENT MEDICAL COLLEGE KOTTAYAM
GANDHINAGAR P.O KOTTAYAM-08

Email ID –
medicalcollegekottayam@gmail.com
Fax- 0481-2597284

Phone Office 0481-2597279
0481-2597284

No.GMCKTM/ 2852/ 2026 -C3 **Date:16-11-2026**

QUOTATION NOTICE

Sealed quotations are invited for the **Purchase of Digital Colorimeter - *4* nos.** (specification attached) **in the Department of Biochemistry, Govt. Medical College, Kottayam.**

The envelopes containing the quotation should bear the subscription "Quotation for the purchase of Digital Colorimeter - 4 nos in the Department of Biochemistry, Govt Medical College, Kottayam and should be addressed to Principal, Govt.Medical College, Kottayam -686 008. Intending tenderers may submit the quotation on their own papers.

Last date and time for receipt of quotation will be 23/11/2026 at 12 O' Clock. Late quotations will not be accepted. The quotations will be opened on 23/11/2026 at 3.00 PM in the presence of such of the tenders or their authorized representatives who may be present at that time. The maximum period required for delivery of the articles should also be mentioned.

Details of the requirements and the conditions governing their supply can be obtained free on request from the Principal, Govt Medical College, Kottayam till 23/11/2026 ,12.00 Noon.

PRINCIPAL

Place: Gandhinagar

Date:16-11-2026

Digitally signed by
Dr Varghese P Punnoose
Date: 16-11-2026
14:08:31

Copy to:

1.Librarian,Central Library (for publishing in website)

2.Notice Board, MCK

3. The HOD, Biochemistry

GOVERNMENT MEDICAL COLLEGE, KOTTAYAM
DEPARTMENT OF BIOCHEMISTRY

COLORIMETER

Specifications:

1. Instrument – Digital Colorimeter.
2. Photometric Range – -0.3 to 2.0 Abs.
3. Spectral Range – 450, 470, 510, 520, 540, 570, 600 & 670 nm filters.
4. Light source – Long life solid state light source
5. Detector – Photo Diode.
6. One-touch auto zero facility must be there.
7. 10 Glass cuvettes suitable for the instrument must be provided along with the instrument.
8. Maximum time between reporting of complaint and attending the same by service personnel must be mentioned.
9. Demonstration should be arranged free of cost.

Dr. SHAMNA BEEGUM.M

DEPARTMENT OF BIOCHEMISTRY
KOTTAYAM
GOVERNMENT MEDICAL COLLEGE